

American National Standard

ANSI/AAMI/IEC 60601-1-2:2007



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,
contact AAMI at (877) 249-8226
or visit www.aami.org.

Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral standard: Electromagnetic compatibility — Requirements and tests



Association for the Advancement
of Medical Instrumentation

The Objectives and Uses of AAMI Standards and Recommended Practices

It is most important that the objectives and potential uses of an AAMI product standard or recommended practice are clearly understood. The objectives of AAMI's technical development program derive from AAMI's overall mission: the advancement of medical instrumentation. Essential to such advancement are (1) a continued increase in the safe and effective application of current technologies to patient care, and (2) the encouragement of new technologies. It is AAMI's view that standards and recommended practices can contribute significantly to the advancement of medical instrumentation, provided that they are drafted with attention to these objectives and provided that arbitrary and restrictive uses are avoided.

A voluntary *standard* for a *medical device* recommends to the manufacturer the information that should be provided with or on the product, basic safety and performance criteria that should be considered in qualifying the device for clinical use, and the measurement techniques that can be used to determine whether the device conforms with the safety and performance criteria and/or to compare the performance characteristics of different products. Some standards emphasize the information that should be provided with the device, including performance characteristics, instructions for use, warnings and precautions, and other data considered important in ensuring the safe and effective use of the device in the clinical environment. Recommending the disclosure of performance characteristics often necessitates the development of specialized test methods to facilitate uniformity in reporting; reaching consensus on these tests can represent a considerable part of committee work. When a drafting committee determines that clinical concerns warrant the establishment of *minimum* safety and performance criteria, referee tests must be provided and the reasons for establishing the criteria must be documented in the rationale.

A *recommended practice* provides guidelines for the use, care, and/or processing of a medical device or system. A recommended practice does not address device performance *per se*, but rather procedures and practices that will help ensure that a device is used safely and effectively and that its performance will be maintained.

Although a device standard is primarily directed to the manufacturer, it may also be of value to the potential purchaser or user of the device as a frame of reference for device evaluation. Similarly, even though a recommended practice is usually oriented towards health care professionals, it may be useful to the manufacturer in better understanding the environment in which a medical device will be used. Also, some recommended practices, while not addressing device performance criteria, provide guidelines to industrial personnel on such subjects as sterilization processing, methods of collecting data to establish safety and efficacy, human engineering, and other processing or evaluation techniques; such guidelines may be useful to health care professionals in understanding industrial practices.

In determining whether an AAMI standard or recommended practice is relevant to the specific needs of a potential user of the document, several important concepts must be recognized:

All AAMI standards and recommended practices are *voluntary* (unless, of course, they are adopted by government regulatory or procurement authorities). The application of a standard or recommended practice is solely within the discretion and professional judgment of the user of the document.

Each AAMI standard or recommended practice reflects the collective expertise of a committee of health care professionals and industrial representatives, whose work has been reviewed nationally (and sometimes internationally). As such, the consensus recommendations embodied in a standard or recommended practice are intended to respond to clinical needs and, ultimately, to help ensure patient safety. A standard or recommended practice is limited, however, in the sense that it responds generally to perceived risks and conditions that may not always be relevant to specific situations. A standard or recommended practice is an important *reference* in responsible decision-making, but it should never *replace* responsible decision-making.

Despite periodic review and revision (at least once every five years), a standard or recommended practice is necessarily a static document applied to a dynamic technology. Therefore, a standards user must carefully review the reasons why the document was initially developed and the specific rationale for each of its provisions. This review will reveal whether the document remains relevant to the specific needs of the user.

Particular care should be taken in applying a product standard to existing devices and equipment, and in applying a recommended practice to current procedures and practices. While observed or potential risks with existing equipment typically form the basis for the safety and performance criteria defined in a standard, professional judgment must be used in applying these criteria to existing equipment. No single source of information will serve to identify a particular product as "unsafe". A voluntary standard can be used as one resource, but the ultimate decision as to product safety and efficacy must take into account the specifics of its utilization and, of course, cost-benefit considerations. Similarly, a recommended practice should be analyzed in the context of the specific needs and resources of the individual institution or firm. Again, the rationale accompanying each AAMI standard and recommended practice is an excellent guide to the reasoning and data underlying its provision.

In summary, a standard or recommended practice is truly useful only when it is used in conjunction with other sources of information and policy guidance and in the context of professional experience and judgment.

INTERPRETATIONS OF AAMI STANDARDS AND RECOMMENDED PRACTICES

Requests for interpretations of AAMI standards and recommended practices must be made in writing, to the Manager for Technical Development. An official interpretation must be approved by letter ballot of the originating committee and subsequently reviewed and approved by the AAMI Standards Board. The interpretation will become official and representation of the Association only upon exhaustion of any appeals and upon publication of notice of interpretation in the "Standards Monitor" section of the *AAMI News*. The Association for the Advancement of Medical Instrumentation disclaims responsibility for any characterization or explanation of a standard or recommended practice which has not been developed and communicated in accordance with this procedure and which is not published, by appropriate notice, as an *official interpretation* in the *AAMI News*.

American National Standard

ANSI/AAMI/IEC 60601-1-2:2007
(Combined revision of
ANSI/AAMI/IEC 60601-1-2:2001 &
ANSI/AAMI/IEC 60601-1-2:2001/Amendment 1:2004)



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,
contact AAMI at (877) 249-8226

Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral standard: Electromagnetic compatibility — Requirements and tests

Approved 15 May 2007 by
Association for the Advancement of Medical Instrumentation

Approved 17 May 2007 by
American National Standards Institute, Inc.

Abstract: Specifies requirements and tests for electromagnetic compatibility of equipment and/or systems and serves as the basis of electromagnetic compatibility requirements and tests in Particular Standards.

Keywords: EMC, electromedical device, electromagnetic interference, medical electrical equipment

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication. Interested parties may obtain current information on all AAMI standards by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,
contact AAMI at (877) 249-8226
or visit www.aami.org.

Published by

Association for the Advancement of Medical Instrumentation
1110 N Glebe Road, Suite 220
Arlington, VA 22201-4795

© 2007 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of IEC, ANSI, and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this draft should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1-57020-287-7

CONTENTS

Glossary of equivalent standards.....	vi
Committee representation.....	viii
Background of AAMI adoption of IEC 60601-1-2:2007.....	ix
FOREWORD.....	x
INTRODUCTION.....	xiii
1 Scope, object, and related standards	1
1.1 * Scope	1
1.2 Object	1
1.3 Related standards.....	1
2 Normative references.....	2
3 Terms and definitions.....	3
4 General requirements	6
4.1 General requirements for ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS.....	6
4.2 * SINGLE FAULT CONDITION for ME EQUIPMENT.....	7
5 Identification, marking and documents	7
5.1 Marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts.....	7
5.2 ACCOMPANYING DOCUMENTS	8
6 ELECTROMAGNETIC COMPATIBILITY	29
6.1 EMISSIONS	29
6.2 IMMUNITY	32
Annex A (informative) General guidance and rationale.....	47
Annex B (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	74
Annex C (informative) Example completion of Table 1 through Table 8	77
Annex D (informative) Guidance in classification according to CISPR 11.....	89
Annex E (informative) Guidance in the application of IEC 60601-1-2 to particular standards	92
Annex F (informative) ELECTROMAGNETIC ENVIRONMENTS	95
Annex G (informative) Guidance for determining if electrical equipment that is not ME EQUIPMENT and that is used in an ME SYSTEM is exempt from the EMC testing requirements of this collateral standard	96
Annex H (informative) Mapping between the elements of the second edition of IEC 60601-1-2 as amended and IEC 60601-1-2:2007.....	98
Bibliography	106
Index of defined terms used in this collateral standard.....	107
Figure 1 – Instructions for completing Table 1 for CISPR 11 ME EQUIPMENT and ME SYSTEMS.....	12
Figure 2 – Instructions for completing Table 1 for CISPR 14 and CISPR 15 ME EQUIPMENT.....	13
Figure 3 – Instructions for completing Table 2.....	16



This is a preview edition of an AAMI guidance document and is not intended for use as a standard. For a complete copy of this AAMI document, contact AAMI at (877) 249-8226 or visit www.aami.org.

Figure 4 – Instructions for completing Table 3 and Table 5 for LIFE-SUPPORTING ME EQUIPMENT and ME SYSTEMS	22
Figure 5 – Instructions for completing Table 4 and Table 6 for ME EQUIPMENT and ME SYSTEMS that are not LIFE-SUPPORTING.....	23
Figure A.1 – Example of cable arrangement for radiated IMMUNITY test.....	72
Figure A.2 – Examples showing maximum dimension for ME EQUIPMENT with one and with two cables.....	73
Figure G.1 – Procedure for determining if electrical equipment that is not ME EQUIPMENT and that is used in an ME SYSTEM is exempt from the EMC testing requirements of this collateral standard	97
Table 1 – Guidance and MANUFACTURER’S declaration – ELECTROMAGNETIC EMISSIONS – for all ME EQUIPMENT and ME SYSTEMS.....	11
Table 2 – Guidance and MANUFACTURER’S declaration – electromagnetic IMMUNITY – for all ME EQUIPMENT and ME SYSTEMS.....	15
Table 3 – Guidance and MANUFACTURER’S declaration – electromagnetic IMMUNITY – for LIFE-SUPPORTING ME EQUIPMENT and ME SYSTEMS.....	18
Table 4 – Guidance and MANUFACTURER’S declaration – electromagnetic IMMUNITY – for ME EQUIPMENT and ME SYSTEMS that are not LIFE-SUPPORTING	19
Table 5 – Recommended separation distances between portable and mobile RF communications equipment and the ME EQUIPMENT or ME SYSTEM – for LIFE-SUPPORTING ME EQUIPMENT and ME SYSTEMS	20
Table 6 – Recommended separation distances between portable and mobile RF communications equipment and the ME EQUIPMENT or ME SYSTEM – for ME EQUIPMENT and ME SYSTEMS that are not LIFE-SUPPORTING	21
Table 7 – Guidance and MANUFACTURER’S declaration – electromagnetic IMMUNITY – for LIFE-SUPPORTING ME EQUIPMENT and ME SYSTEMS that are specified for use only in a shielded location	27
Table 8 – Guidance and MANUFACTURER’S declaration – electromagnetic IMMUNITY – for ME EQUIPMENT and ME SYSTEMS that are not LIFE-SUPPORTING and are specified for use only in a shielded location	28
Table 9 – Modulation frequency, PHYSIOLOGICAL SIMULATION FREQUENCY, and OPERATING FREQUENCY	36
Table 10 – IMMUNITY TEST LEVELS for voltage dips	44
Table 11 – IMMUNITY TEST LEVEL for voltage interruption.....	44
Table B.1 – Marking on the outside of ME EQUIPMENT, ME SYSTEMS, or their parts	74
Table B.2 – ACCOMPANYING DOCUMENTS, instructions for use.....	75
Table B.3 – ACCOMPANYING DOCUMENTS, technical description	76
Table C.1 – Example (1) of completed Table 1	77
Table C.2 – Example (2) of completed Table 1	78
Table C.3 – Example (3) of completed Table 1	79
Table C.4 – Example of completed Table 2.....	80
Table C.5 – Example (1) test, IMMUNITY, and COMPLIANCE LEVELS	81
Table C.6 – Example of completed Table 3.....	82
Table C.7 – Example of completed Table 5.....	83
Table C.8 – Example of completed Table 4.....	84

Table C.9 – Example of completed Table 6.....	85
Table C.10 – Example (2) test, IMMUNITY, and COMPLIANCE LEVELS	85
Table C.11 – Example of completed Table 7.....	86
Table C.12 – Example (3) test, IMMUNITY, and COMPLIANCE LEVELS	87
Table C.13 – Example of completed Table 8.....	88
Table F.1 – ELECTROMAGNETIC ENVIRONMENTS	95
Table H.1 – Mapping between the elements of the second edition of IEC 60601-1-2 as amended and IEC 60601-1-2:2007	98



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,
contact AAMI at (877) 249-8226
or visit www.aami.org.

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. For each International Standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the International Standard. NOTE: Documents are sorted by international designation.

Other normatively referenced International Standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1:2005	ANSI/AAMI ES60601-1:2005	Major technical variations
IEC 60601-1-2:2007	ANSI/AAMI/IEC 60601-1-2:2007	Identical
IEC 60601-2-2:2006	ANSI/AAMI/IEC 60601-2-2:2006	Identical
IEC 60601-2-4:2002	ANSI/AAMI DF80:2003	Major technical variations
IEC 60601-2-19:1990 and Amendment 1:1996	ANSI/AAMI II36:2004	Major technical variations
IEC 60601-2-20:1990 and Amendment 1:1996	ANSI/AAMI II51:2004	Major technical variations
IEC 60601-2-21:1994 and Amendment 1:1996	ANSI/AAMI/IEC 60601-2-21 and Amendment 1:2000 (consolidated texts)	Identical
IEC 60601-2-24:1998	ANSI/AAMI ID26:2004	Major technical variations
IEC 60601-2-50:2001	ANSI/AAMI/IEC 60601-2-50:2006	Identical
IEC/TR 60878:2003	ANSI/AAMI/IEC TIR60878:2003	Identical
IEC/TR 62296:2003	ANSI/AAMI/IEC TIR62296:2003	Identical
IEC 62304:2006	ANSI/AAMI/IEC 62304:2006	Identical
IEC/TR 62348:2006	ANSI/AAMI/IEC TIR62348:2006	Identical
ISO 5840:2005	ANSI/AAMI/ISO 5840:2005	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001/(R)2004	Identical
ISO 7199:1996	ANSI/AAMI/ISO 7199:1996/(R)2002	Identical
ISO 8637:2004	ANSI/AAMI RD16:2007	Major technical variations
ISO 8638:2004	ANSI/AAMI RD17:2007	Major technical variations
ISO 10993-1:2003	ANSI/AAMI/ISO 10993-1:2003	Identical
ISO 10993-2:2006	ANSI/AAMI/ISO 10993-2:2006	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003	Identical
ISO 10993-4:2002 and Amendment 1:2006	ANSI/AAMI/ISO 10993-4:2002 and Amendment 1:2006	Identical
ISO 10993-5:1999	ANSI/AAMI/ISO 10993-5:1999	Identical
ISO 10993-6:2007	ANSI/AAMI/ISO 10993-6:2007	Identical
ISO 10993-7:1995	ANSI/AAMI/ISO 10993-7:1995/(R)2001	Identical
ISO 10993-9:1999	ANSI/AAMI/ISO 10993-9:1999/(R)2005	Identical
ISO 10993-10:2002 and Amendment 1:2006	ANSI/AAMI BE78:2002 ANSI/AAMI BE78:2002/A1:2006	Minor technical variations Identical
ISO 10993-11:2006	ANSI/AAMI/ISO 10993-11:2006	Identical
ISO 10993-12:2002	ANSI/AAMI/ISO 10993-12:2002	Identical
ISO 10993-13:1998	ANSI/AAMI/ISO 10993-13:1999/(R)2004	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001/(R)2006	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000/(R)2006	Identical

International designation	U.S. designation	Equivalency
ISO 10993-16:1997	ANSI/AAMI/ISO 10993-16:1997/(R)2003	Identical
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002	Identical
ISO 10993-18:2005	ANSI/AAMI BE83:2006	Major technical variations
ISO/TS 10993-19:2006	ANSI/AAMI/ISO TIR10993-19:2006	Identical
ISO/TS 10993-20:2006	ANSI/AAMI/ISO TIR10993-20:2006	Identical
ISO 11135:1994	ANSI/AAMI/ISO 11135:1994	Identical
ISO 11137-1:2006	ANSI/AAMI/ISO 11137-1:2006	Identical
ISO 11137-2:2006 (2006-08-01 corrected version)	ANSI/AAMI/ISO 11137-2:2006	Identical
ISO 11137-3:2006	ANSI/AAMI/ISO 11137-3:2006	Identical
ISO 11138-1: 2006	ANSI/AAMI/ISO 11138-1:2006	Identical
ISO 11138-2: 2006	ANSI/AAMI/ISO 11138-2:2006	Identical
ISO 11138-3: 2006	ANSI/AAMI/ISO 11138-3:2006	Identical
ISO 11138-4: 2006	ANSI/AAMI/ISO 11138-4:2006	Identical
ISO 11138-5: 2006	ANSI/AAMI/ISO 11138-5:2006	Identical
ISO/TS 11139:2006	ANSI/AAMI/ISO 11139:2006	Identical
ISO 11140-1:2005	ANSI/AAMI/ISO 11140-1:2005	Identical
ISO 11140-3:2007	ANSI/AAMI/ISO 11140-3:2007	Identical
ISO 11140-4:2007	ANSI/AAMI/ISO 11140-4:2007	Identical
ISO 11140-5:2007	ANSI/AAMI/ISO 11140-5:2007	Identical
ISO 11607-1:2006	ANSI/AAMI/ISO 11607-1:2006	Identical
ISO 11607-2:2006	ANSI/AAMI/ISO 11607-2:2006	Identical
ISO 11737-1: 2006	ANSI/AAMI/ISO 11737-1:2006	Identical
ISO 11737-2:1998	ANSI/AAMI/ISO 11737-2:1998	Identical
ISO 11737-3:2004	ANSI/AAMI/ISO 11737-3:2004	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003	Identical
ISO 14155-1:2003	ANSI/AAMI/ISO 14155-1:2003	Identical
ISO 14155-2:2003	ANSI/AAMI/ISO 14155-2:2003	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998	Identical
ISO 14161:2000	ANSI/AAMI/ISO 14161:2000	Identical
ISO 14937:2000	ANSI/AAMI/ISO 14937:2000	Identical
ISO/TR 14969:2004	ANSI/AAMI/ISO TIR14969:2004	Identical
ISO 14971:2007	ANSI/AAMI/ISO 14971:2007	Identical
ISO 15223-1:2007	ANSI/AAMI/ISO 15223-1:2007	Identical
ISO 15225:2000 and A1:2004	ANSI/AAMI/ISO 15225:2000/(R)2006 and A1:2004/(R)2006	Identical
ISO 15674:2001	ANSI/AAMI/ISO 15674:2001	Identical
ISO 15675:2001	ANSI/AAMI/ISO 15675:2001	Identical
ISO 15882:2003	ANSI/AAMI/ISO 15882:2003	Identical
ISO/TR 16142:2006	ANSI/AAMI/ISO TIR16142:2006	Identical
ISO 17664:2004	ANSI/AAMI ST81:2004	Major technical variations
ISO 17665-1:2006	ANSI/AAMI/ISO 17665-1:2006	Identical
ISO 18472:2006	ANSI/AAMI/ISO 18472:2006	Identical
ISO/TS 19218:2005	ANSI/AAMI/ISO 19218:2005	Identical
ISO 25539-1:2003 and A1:2005	ANSI/AAMI/ISO 25539-1:2003 and A1:2005	Identical

Committee representation

Association for the Advancement of Medical Instrumentation

Electromagnetic Compatibility Committee

This standard was developed by the Electromagnetic Compatibility Committee of the Association for the Advancement of Medical Instrumentation. Committee approval of this standard does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Electromagnetic Compatibility Committee** had the following members:

Cochairs: Don Heirman
Dara McLain

Members: Eric V. Anderson, Philips Medical Systems
Art Augustine, ECRI
Alan S. Berson, PhD, Bioresearch Funding Group
Steve Cantwell, Spacelabs Medical Inc.
Jimmy Cheng, Respironics Inc.
Yadin David, PhD, Texas Childrens Hospital
Joseph F. Dyro, PhD, East Setauket, NY
Wallace R. Elliott, University of Vermont
James J. Greco, Medapprove Inc.
Michael Gusel, Datascope Corp.
Rickey L. Hampton, Partners Healthcare Systems
Donald N. Heirman, University of Oklahoma
Veronica Ivans, Medtronic Inc.
Robert S. Jenkins, Welch Allyn Inc.
Donald Koller, GE Healthcare
Bernie Liebler, Advanced Medical Technology Association (AdvaMed)
John Manarik, Baxter Healthcare Corp.
Joseph P. McClain, PhD, Walter Reed Army Medical Center
Bernard N. Segal, PhD, Jewish General Hospital
Jeffrey Silberberg, U.S. Food and Drug Administration
Ray P. Silkaitis, PhD, Hospira Inc.
James Spicer, Abbott Laboratories
James D. Stewardson, Brighton, CO
Joshua E. Tsitlik, PhD, Washington Hospital Center
David Tyler, Gambro Technical Services
Stephen Vastagh, National Electrical Manufacturers Association (NEMA)
Richard J. Wessels, Guidant Corp.

Alternates: Lawrence Baylis, Medtronic Inc.
Scott Hornbrook, Spacelabs Medical Inc.
Jamie Jones, Abbott Laboratories
Joshua Kim, Welch Allyn Inc
Steve Kreinick, Abbott Laboratories
Dara Mclain, Philips Medical Systems
Gopal Mohanty, Baxter Healthcare Corp.
David Osborn, Philips Medical Systems
Ashwin Patel, Hospira Inc.
Donald M. Witters, Jr., U.S. Food and Drug Administration

NOTE—Participation by federal agency representatives in the development of this recommended practice does not constitute endorsement by the federal government or any of its agencies.

Background of AAMI adoption of IEC 60601-1-2, Third edition, 2007-04

As indicated in the foreword to the main body of this document (page x), the International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The United States is one of the IEC members that took an active role in the development of this standard.

International Standard IEC 60601-1-2 was developed by Maintenance Team (MT) 23, Electromagnetic Compatibility, of Subcommittee (SC) 62A, Common aspects of electrical equipment used in medical practice, to provide minimum safety requirements that will help assure a reasonable level of clinical efficacy and patient safety.

U.S. participation in IEC/SC 62A/MT 23 is organized through the U.S. Technical Advisory Group for IEC/SC 62A, administered by the Advanced Medical Technology Association (AdvaMed) on behalf of the United States National Committee. AAMI administers the International Secretariat for IEC/SC 62A on behalf of the United States, and U.S. experts made a considerable contribution to this standard.

AAMI encourages its committees to harmonize their work with international standards as much as possible. The AAMI Electromagnetic Compatibility Committee held joint meetings with the U.S. Technical Advisory Group for IEC/SC 62A to formulate the U.S. position and comments while the document was being developed. This close collaboration helped gain widespread consensus on the document. As the U.S. Technical Advisory Group for IEC/SC 62A, AdvaMed granted AAMI permission for an identical adoption of IEC 60601-1-2:2007 (3rd ed.) as a combined revision of the American National Standard ANSI/AAMI/IEC 60601-1-2:2001 and its Amendment, ANSI/AAMI/IEC 60601-1-2:2001/A1:2004.

IEC 60601-1-2 is a collateral standard to IEC 60601-1. The AAMI adoption of IEC 60601-1:2005 as an American National Standard included U.S. deviations and was subsequently designated ANSI/AAMI ES60601-1:2005. Therefore, ANSI/AAMI/IEC 60601-1-2:2007 is also a collateral standard to ANSI/AAMI ES60601-1:2005.

The concepts incorporated into this standard should not be considered inflexible or static. This standard, like any other, must be reviewed and updated periodically to assimilate progressive technological developments. To remain relevant, it must be modified as technological advances are made and as new data come to light.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Standards Department, AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

NOTE 1—This background does not contain provisions of the AAMI/IEC standard, *Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral standard: Electromagnetic compatibility — Requirements and tests* (AAMI/IEC 60601-1-2:2007), but it does provide important information about the development and intended use of the document.

NOTE 2—Beginning with the text on page vii, this American National Standard is identical to IEC 60601-1-2, Third edition, 2007-04.

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental, and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-1-2 has been prepared by IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition constitutes a collateral standard to IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for safety and essential performance* hereafter referred to as the general standard.

This document cancels and replaces the second edition of IEC 60601-1-2, and constitutes a technical revision.

This edition of IEC 60601-1-2 was revised to structurally align it with the 2005 edition of IEC 60601-1 and to implement the decision of IEC subcommittee 62A that the clause numbering structure of collateral standards written to IEC 60601-1:2005 would adhere to the form specified in ISO/IEC Directives, Part 2:2004. The principle technical changes are in Clause 4, which now recognizes that there is a general requirement for a risk management process in IEC 60601-1:2005.

The text of this standard is based on the following documents:

FDIS	Report on voting
62A/560/FDIS	62A/567/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. alarm systems)

In this collateral standard, the following print types are used:

- requirements and definitions: roman type.
- *test specifications: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

NOTE Defined terms are not printed in SMALL CAPITALS in Table 1 through Table 8, in the tables in Annex C and in statements required to appear in the technical description or instructions for use because they are intended for the OPERATOR or RESPONSIBLE ORGANIZATION, who may not be familiar with the defined terms of IEC 60601 standards.

In referring to the structure of this standard, the term

- “clause” means one of the six numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 6 includes 6.1, 6.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 6.1, 6.2, and 6.2.1 are all subclauses of Clause 6).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses, items, and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (*).

A list of all parts of the IEC 60601 series, under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document, contact AAMI at (877) 249-8226 or visit www.aami.org.

INTRODUCTION

The need for establishing specific ELECTROMAGNETIC COMPATIBILITY standards for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS is well recognized.

In particular, the existence of ELECTROMAGNETIC EMISSION standards is essential for the protection of:

- safety services;
- other MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS;
- electrical equipment that is not ME EQUIPMENT (e.g. computers);
- telecommunications (e.g. radio/TV, telephone, radio-navigation).

Of even more importance, the existence of ELECTROMAGNETIC IMMUNITY standards is essential to assure safety of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS. ELECTROMAGNETIC COMPATIBILITY (see Definition 3.4) differs from other aspects of safety covered by IEC 60601-1 because the electromagnetic phenomena exist, with varying degrees of severity, in the normal use environment of all MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS and by definition the equipment must “perform satisfactorily” within its intended environment in order to establish ELECTROMAGNETIC COMPATIBILITY. This means that the conventional single fault approach to safety is not appropriate for application to ELECTROMAGNETIC COMPATIBILITY standards. The ELECTROMAGNETIC DISTURBANCE environment can be compared to ambient temperature, humidity, and atmospheric pressure. MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS may experience environmental conditions within the expected range at any time, and for extended periods of time. As with atmospheric pressure and humidity, the OPERATOR of the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM may not be aware of ambient levels on a continuous basis. The IMMUNITY TEST LEVELS specified in this collateral standard (IEC 60601 TEST LEVELS) represent the range found in the general medical use environment. Therefore, under these conditions, the performance of the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM would also be expected to be normal.

MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS are used in the practice of medicine because they provide needed FUNCTIONS. If MEDICAL ELECTRICAL EQUIPMENT or a MEDICAL ELECTRICAL SYSTEM does not provide its needed FUNCTION, because of a lack of IMMUNITY to events expected in the normal use environment, this interferes with the practice of medicine and cannot be considered an acceptable situation.

This edition recognizes that there is a shared responsibility between MANUFACTURERS, RESPONSIBLE ORGANIZATIONS, and OPERATORS to ensure that MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS are designed and operated as intended. The MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM MANUFACTURER’S responsibility is to design and manufacture to meet the requirements of this collateral standard and to disclose information to the RESPONSIBLE ORGANIZATION or OPERATOR so that a compatible ELECTROMAGNETIC ENVIRONMENT can be maintained in order that the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM will perform as intended.

Because the practice of medicine involves many specialties, there will by necessity be MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS that are designed to perform a variety of FUNCTIONS. Some FUNCTIONS involve, for example, measurement of signals from a PATIENT that are of very low levels when compared to ELECTROMAGNETIC NOISE levels that can be coupled into MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS during the ELECTROMAGNETIC IMMUNITY testing specified in this collateral standard. Because of the proven benefits of many such MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, this collateral standard allows the IMMUNITY TEST LEVELS to be lowered, provided there is sufficient justification based on physical, technological, or physiological limitations. In this case, the MANUFACTURER is required to disclose the levels at which the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM meets the performance requirements of this collateral standard and to specify the characteristics of the ELECTROMAGNETIC use environment and how this environment is established, in which the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM will perform as intended.

This collateral standard also recognizes that for certain environments, higher IMMUNITY LEVELS may be required. Research necessary to determine how to identify the environments that may require higher IMMUNITY LEVELS, as well as what the levels should be, is in progress.

Finally, this collateral standard recognizes that for LIFE-SUPPORTING MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, higher levels of IMMUNITY are necessary in order to establish a broader safety margin, even for use in the general medical use environment. Therefore, this collateral standard specifies additional requirements for LIFE-SUPPORTING MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS.

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

This collateral standard is based on existing IEC standards prepared by subcommittee 62A, technical committee 77 (electromagnetic compatibility between electrical equipment including networks) and CISPR (International special committee on radio interference).

The ELECTROMAGNETIC COMPATIBILITY requirements specified by this collateral standard are generally applicable to MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS as defined in 3.63 and 3.64 in the general standard. For certain types of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, these requirements may need to be modified by the special requirements of a particular standard. Writers of particular standards are encouraged to refer to Annex E for guidance in the application of this collateral standard.

Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral standard: Electromagnetic compatibility — Requirements and tests

1 Scope, object, and related standards

1.1 * Scope

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, hereafter referred to as ME EQUIPMENT and ME SYSTEMS.

This collateral standard applies to ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS.

1.2 Object

The object of this collateral standard is to specify general requirements and tests for ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS. They are in addition to the requirements of the general standard and serve as the basis for particular standards.

1.3 Related standards

1.3.1 IEC 60601-1

For ME EQUIPMENT and ME SYSTEMS, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1 alone;
- "this collateral standard" designates IEC 60601-1-2 alone;
- "this standard" designates the combination of the general standard and this collateral standard.

1.3.2 Particular standards

A requirement in a particular standard takes priority over the corresponding requirement in this collateral standard.