

**American
National
Standard**

ANSI/AAMI DF80:2003

**Medical electrical equipment—
Part 2-4: Particular
requirements for the safety of
cardiac defibrillators
(including automated
external defibrillators)**

The Objectives and Uses of AAMI Standards and Recommended Practices

It is most important that the objectives and potential uses of an AAMI product standard or recommended practice are clearly understood. The objectives of AAMI's technical development program derive from AAMI's overall mission: the advancement of medical instrumentation. Essential to such advancement are (1) a continued increase in the safe and effective application of current technologies to patient care, and (2) the encouragement of new technologies. It is AAMI's view that standards and recommended practices can contribute significantly to the advancement of medical instrumentation, provided that they are drafted with attention to these objectives and provided that arbitrary and restrictive uses are avoided.

A voluntary *standard* for a *medical device* recommends to the manufacturer the information that should be provided with or on the product, basic safety and performance criteria that should be considered in qualifying the device for clinical use, and the measurement techniques that can be used to determine whether the device conforms with the safety and performance criteria and/or to compare the performance characteristics of different products. Some standards emphasize the information that should be provided with the device, including performance characteristics, instructions for use, warnings and precautions, and other data considered important in ensuring the safe and effective use of the device in the clinical environment. Recommending the disclosure of performance characteristics often necessitates the development of specialized test methods to facilitate uniformity in reporting; reaching consensus on these tests can represent a considerable part of committee work. When a drafting committee determines that clinical concerns warrant the establishment of *minimum* safety and performance criteria, referee tests must be provided and the reasons for establishing the criteria must be documented in the rationale.

A *recommended practice* provides guidelines for the use, care, and/or processing of a medical device or system. A recommended practice does not address device performance *per se*, but rather procedures and practices that will help ensure that a device is used safely and effectively and that its performance will be maintained.

Although a device standard is primarily directed to the manufacturer, it may also be of value to the potential purchaser or user of the device as a fume of reference for device evaluation. Similarly, even though a recommended practice is usually oriented towards health care professionals, it may be useful to the manufacturer in better understanding the environment in which a medical device will be used. Also, some recommended practices, while not addressing device performance criteria, provide guidelines to industrial personnel on such subjects as sterilization processing, methods of collecting data to establish safety and efficacy, human engineering, and other processing or evaluation techniques; such guidelines may be useful to health care professionals in understanding industrial practices.

In determining whether an AAMI standard or recommended practice is relevant to the specific needs of a potential user of the document, several important concepts must be recognized:

All AAMI standards and recommended practices are *voluntary* (unless, of course, they are adopted by government regulatory or procurement authorities). The application of a standard or recommended practice is solely within the discretion and professional judgment of the user of the document.

Each AAMI standard or recommended practice reflects the collective expertise of a committee of health care professionals and industrial representatives, whose work has been reviewed nationally (and sometimes internationally). As such, the consensus recommendations embodied in a standard or recommended practice are intended to respond to clinical needs and, ultimately, to help ensure patient safety. A standard or recommended practice is limited, however, in the sense that it responds generally to perceived risks and conditions that may not always be relevant to specific situations. A standard or recommended practice is an important *reference* in responsible decision-making, but it should never *replace* responsible decisionmaking.

Despite periodic review and revision (at least once every five years), a standard or recommended practice is necessarily a static document applied to a dynamic technology. Therefore, a standards user must carefully review the reasons why the document was initially developed and the specific rationale for each of its provisions. This review will reveal whether the document remains relevant to the specific needs of the user.

Particular care should be taken in applying a product standard to existing devices and equipment, and in applying a recommended practice to current procedures and practices. While observed or potential risks with existing equipment typically form the basis for the safety and performance criteria defined in a standard, professional judgment must be used in applying these criteria to existing equipment. No single source of information will serve to identify a particular product as "unsafe". A voluntary standard can be used as one resource, but the ultimate decision as to product safety and efficacy must take into account the specifics of its utilization and, of course, cost-benefit considerations. Similarly, a recommended practice should be analyzed in the context of the specific needs and resources of the individual institution or firm. Again, the rationale accompanying each AAMI standard and recommended practice is an excellent guide to the reasoning and data underlying its provision.

In summary, a standard or recommended practice is truly useful only when it is used in conjunction with other sources of information and policy guidance and in the context of professional experience and judgment.

INTERPRETATIONS OF AAMI STANDARDS AND RECOMMENDED PRACTICES

Requests for interpretations of AAMI standards and recommended practices must be made in writing, to the Manager for Technical Development. An official interpretation must be approved by letter ballot of the originating committee and subsequently reviewed and approved by the AAMI Standards Board. The interpretation will become official and representation of the Association only upon exhaustion of any appeals and upon publication of notice of interpretation in the "Standards Monitor" section of the *AAMI News*. The Association for the Advancement of Medical Instrumentation disclaims responsibility for any characterization or explanation of a standard or recommended practice which has not been developed and communicated in accordance with this procedure and which is not published, by appropriate notice, as an *official interpretation* in the *AAMI News*.



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,
contact AAMI at (800) 332-2264, ext. 217
or visit www.aami.org.

Medical electrical equipment—Part 2-4: Particular requirements for the safety of cardiac defibrillators (including automated external defibrillators)

Developed by
Association for the Advancement of Medical Instrumentation

Approved 23 October 2003 by
American National Standards Institute, Inc.

Abstract: This standard specifies requirements for the safety of medical electrical equipment intended to defibrillate the heart by an electrical pulse via electrodes applied either to the patient's skin (external electrodes) or to the exposed heart (internal electrodes). This standard does not apply to implantable defibrillators, remote control defibrillators, or separate cardiac monitors.

Keywords: manual defibrillator, automatic external defibrillator, AED, electromedical equipment, cardiac

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of publication. Interested parties may obtain current information on all AAMI standards by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

PREVIEW COPY

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,
contact AAMI at (800) 332-2264, ext. 217
or visit www.aami.org.

Published by

Association for the Advancement of Medical Instrumentation
1110 N Glebe Road, Suite 220
Arlington, VA 22201-4795

© 2004 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of IEC, ANSI, and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this draft should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, et seq.) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795. Phone: (703) 525-4890; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1-57020-210-9

Contents

	Page
Glossary of equivalent standards	vi
Committee representation	viii
Background.....	ix
FOREWORD.....	x
INTRODUCTION.....	xii
SECTION ONE – GENERAL	
1 Scope and object	1
2 Terminology and definitions	2
4 General requirements for tests	5
*5 Classification.....	5
6 Identification, marking, and documents	5
SECTION TWO – ENVIRONMENTAL CONDITIONS	
10 Environmental conditions.....	10
SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS	
14 Requirements related to classification	10
*17 Separation.....	10
19 Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS	11
*20 Dielectric strength	13
SECTION FOUR – PROTECTION AGAINST MECHANICAL HAZARDS	
SECTION FIVE – PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION	
*36 Electromagnetic compatibility (EMC).....	15
SECTION SIX– PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES	
SECTION SEVEN – PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS	
*42 Excessive temperatures.....	18
44 Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization, disinfection, and compatibility	18
46 Human errors	19

SECTION EIGHT – ACCURACY OF OPERATING DATA
AND PROTECTION AGAINST HAZARDOUS OUTPUT

*50 Accuracy of operating data20
51 Protection against hazardous output20

SECTION NINE – ABNORMAL OPERATION AND FAULT CONDITIONS;
ENVIRONMENTAL TESTS

52 Abnormal operation and fault conditions21

SECTION TEN – CONSTRUCTIONAL REQUIREMENTS

*56 Components and general assembly21
57 MAINS PARTS, components and layout23

SECTION 101 – ADDITIONAL REQUIREMENTS RELATING TO SAFETY

*101 Charging time23
102 Internal electrical power source26
*103 Endurance28
*104 Synchronizer28
*105 Recovery of the MONITOR/ECG INPUT after defibrillation28
*106 Disturbance to the monitor from charging or internal discharging30
*107 Defibrillator electrodes (U.S.)50
*108 External pacing (U.S.)52

Annexes

Appendix L References – Publications mentioned in this Standard37
Annex AA (informative) General guidance and rationale38
Annex BB (informative) AUTOMATED EXTERNAL DEFIBRILLATORS: background and rationale48
Annex CC (informative) General guidance and rationale for additional U.S. requirements (U.S.)58
Annex DD (informative) Historical background on defibrillation waveforms (U.S.)61
Annex EE (informative) Cited references in Annexes CC and DD (U.S.)66

Tables

101 Rhythm recognition detector categories9
US.DD.1 Specifications for damped sinusoidal output waveforms64
US.DD.2 Specifications for monophasic truncated exponential output waveforms65

Figures

101 Dynamic test for limitation of energy from different parts of the EQUIPMENT31
102 Test apparatus for flexible cords and their anchorages32

103	Arrangement for test of recovery after defibrillation	32
104	Arrangement of monitoring electrodes on sponge	33
105	Arrangement for recovery test after defibrillation	33
106	Arrangement for test of disturbance from charging and internal discharging	34
107	Isolation of DEFIBRILLATOR circuit from separate monitoring electrodes	34
108	Examples of cord anchorages that require testing	35
US.1	Test circuit for offset instability/internal noise determination	56
US.2	Test circuit for defibrillator overload test of pacing output circuitry	57
US.DD.1	Damped sinusoidal waveform parameters	64
US.DD.2	Truncated exponential waveform parameters	65



This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision.

For a complete copy of this AAMI document,
 contact AAMI at (800) 332-2264, ext. 217
 or visit www.aami.org.

Glossary of equivalent standards

International standards adopted in the United States may include normative references to other international standards. For each international standard that has been adopted by AAMI (and ANSI), the table below gives the corresponding U.S. designation and level of equivalency to the international standard. (Note: Documents are sorted by international designation.)

Other normatively referenced international standards may be under consideration for U.S. adoption by AAMI; therefore, this list should not be considered exhaustive.

International designation	U.S. designation	Equivalency
IEC 60601-1-2:2001	ANSI/AAMI/IEC 60601-1-2:2001	Identical
IEC 60601-2-04:2002	ANSI/AAMI DF80:2003	Major technical variations
IEC 60601-2-21:1994 and Amendment 1:1996	ANSI/AAMI/IEC 60601-2-21 & Amendment 1:2000 (consolidated texts)	Identical
IEC 60601-2-24:1998	ANSI/AAMI JD26:1998	Major technical variations
IEC TR 60878:2003	ANSI/AAMI/IEC TR 60878:2003	Identical
IEC TR 62296:2003	ANSI/AAMI/IEC TR 62296:2003	Identical
ISO 5840:1996	ANSI/AAMI/ISO 5840:1996	Identical
ISO 7198:1998	ANSI/AAMI/ISO 7198:1998/2001	Identical
ISO 7199:1996	ANSI/AAMI/ISO 7199:1996/(R)2002	Identical
ISO 10993-1:2003	ANSI/AAMI/ISO 10993-1:2003	Identical
ISO 10993-2:1992	ANSI/AAMI/ISO 10993-2:1993/(R)2001	Identical
ISO 10993-3:2003	ANSI/AAMI/ISO 10993-3:2003	Identical
ISO 10993-4:2002	ANSI/AAMI/ISO 10993-4:2002	Identical
ISO 10993-5:1999	ANSI/AAMI/ISO 10993-5:1999	Identical
ISO 10993-6:1994	ANSI/AAMI/ISO 10993-6:1995/(R)2001	Identical
ISO 10993-7:1995	ANSI/AAMI/ISO 10993-7:1995/(R)2001	Identical
ISO 10993-8:2000	ANSI/AAMI/ISO 10993-8:2000	Identical
ISO 10993-9:1999	ANSI/AAMI/ISO 10993-9:1999	Identical
ISO 10993-10:2002	ANSI/AAMI BE78:2002	Minor technical variations
ISO 10993-11:1993	ANSI/AAMI 10993-11:1993	Minor technical variations
ISO 10993-12:2002	ANSI/AAMI/ISO 10993-12:2002	Identical
ISO 10993-13:1998	ANSI/AAMI/ISO 10993-13:1999	Identical
ISO 10993-14:2001	ANSI/AAMI/ISO 10993-14:2001	Identical
ISO 10993-15:2000	ANSI/AAMI/ISO 10993-15:2000	Identical
ISO 10993-16:1997	ANSI/AAMI/ISO 10993-16:1997/(R)2003	Identical

International designation	U.S. designation	Equivalency
ISO 10993-17:2002	ANSI/AAMI/ISO 10993-17:2002	Identical
ISO 11134:1994	ANSI/AAMI/ISO 11134:1993	Identical
ISO 11135:1994	ANSI/AAMI/ISO 11135:1994	Identical
ISO 11137:1995 and Amdt 1:2001	ANSI/AAMI/ISO 11137:1994 and A1:2002	Identical
ISO 11138-1:1994	ANSI/AAMI ST59:1999	Major technical variations
ISO 11138-2:1994	ANSI/AAMI ST21:1999	Major technical variations
ISO 11138-3:1995	ANSI/AAMI ST19:1999	Major technical variations
ISO TS 11139:2001	ANSI/AAMI/ISO 11139:2002	Identical
ISO 11140-1:1995 and Technical Corrigendum 1:1998	ANSI/AAMI ST60:1996	Major technical variations
ISO 11607:2003	ANSI/AAMI/ISO 11607:2000	Identical
ISO 11737-1:1995	ANSI/AAMI/ISO 11737-1:1995	Identical
ISO 11737-2:1998	ANSI/AAMI/ISO 11737-2:1998	Identical
ISO TR 13409:1996	AAMI/ISO TIR13409:1996	Identical
ISO 13485:2003	ANSI/AAMI/ISO 13485:2003	Identical
ISO 13488:1996	ANSI/AAMI/ISO 13488:1996	Identical
ISO 14155-1:2003	ANSI/AAMI/ISO 14155-1:2003	Identical
ISO 14155-2:2003	ANSI/AAMI/ISO 14155-2:2003	Identical
ISO 14160:1998	ANSI/AAMI/ISO 14160:1998	Identical
ISO 14161: 2000	ANSI/AAMI/ISO 14161:2000	Identical
ISO 14937:2000	ANSI/AAMI/ISO 14937:2000	Identical
ISO 14969:1999	ANSI/AAMI/ISO 14969:1999	Identical
ISO 14971:2000 and A1:2003	ANSI/AAMI/ISO 14971:2000 and A1:2003	Identical
ISO 15223:2000	ANSI/AAMI/ISO 15223:2000	Identical
ISO 15223/A1:2002	ANSI/AAMI/ISO 15223:2000/A1:2001	Identical
ISO 15225:2000	ANSI/AAMI/ISO 15225:2000	Identical
ISO 15674:2001	ANSI/AAMI/ISO 15674:2001	Identical
ISO 15675:2001	ANSI/AAMI/ISO 15675:2001	Identical
ISO TS 15843:2000	ANSI/AAMI/ISO TIR15843:2000	Identical
ISO TR 15844:1998	AAMI/ISO TIR15844:1998	Identical
ISO TR 16142:1999	ANSI/AAMI/ISO TIR16142:2000	Identical
ISO 25539-1:2003	ANSI/AAMI/ISO 25539-1:2003	Identical

Committee representation

Association for the Advancement of Medical Instrumentation

Defibrillator Committee

This standard was developed by the Defibrillator Committee of the Association for the Advancement of Medical Instrumentation. Committee approval of the standard does not necessarily imply that all committee members voted for its approval.

At the time this document was published, the **AAMI Defibrillator Committee** had the following members:

Cochairs: Richard Kerber, MD
Michael Willingham

Members: John Anderson, University of Ulster
Robert William Bain, Prince George's Hospital Center
Carole C. Carey, RN, U. S. Food and Drug Administration
Hatim M. Carim, PhD, 3M Healthcare
Andrew C. Clifford, SGS Medical Devices
Regis DeSilva, MD, Harvard Medical School
Peter D. Gadsby, Tyco Healthcare/Ludlow
Leslie A. Geddes, PhD, Purdue University
Janice M. Jenkins, PhD, University of Michigan College of Engineering
Gideon Kantor, PhD
Richard Kerber, MD, University of Iowa Healthcare
Jim Miller, Philips Medical Systems
Carl A. Pantiskas, Spacelabs Medical
Cameron G. Rouns, Ballard Medical Products
David Schlageter, G. E. Marquette Medical Systems
William J. Smirles, Heartsine Technologies, Inc.
W. A. Tacker, Jr., MD, PhD, Purdue University
Kok-Swang Tan, PhD, Medical Devices Bureau Health
Michael D. Willingham, Medtronic-Physio Control

Alternates: Donald Eugene Brodnick, G.E. Medical Systems
Fred W. Chapman, PhD, Medtronic Physio-Control
Stacy E. Gehman, BS, Philips Medical Systems/Heartstream
Shen Luo, PhD, Spacelabs Medical

NOTE—Participation by federal agency representatives in the development of this standard does not constitute endorsement by the federal government or any of its agencies.

Background

This standard was developed by the AAMI Defibrillator Committee. The objective of this standard is to specify requirements for the safety of medical electrical equipment intended to defibrillate the heart by an electrical pulse via electrodes applied either to the patient's skin or to the exposed heart.

This is a combined revision of ANSI/AAMI DF2:1996, *Cardiac defibrillator devices* and ANSI/AAMI DF39:1993, *Automatic external defibrillators and remote-control defibrillators*. During the course of putting this document together, the AAMI Defibrillator Committee considered identical adoption of IEC 60601-2-4:2002, *Medical electrical equipment—Part 2-4: Particular requirements for the safety of cardiac defibrillator devices*, which was based in large part on the previous editions of the AAMI standards noted above. In the course of reviewing the IEC standard for U.S. adoption, the committee decided that although the IEC document addressed most of the issues which the committee members felt were important, there were a few areas covered in DF2 and DF39 that had been omitted by IEC. This AAMI standard, therefore, includes all of the requirements from IEC 60601-2-4:2002 as well as some additional requirements and informative text that apply only to the AAMI standard.

The additional requirements and other informative materials that relate only to the American National Standard consist of sections 107 and 108, and informative annexes CC, DD, and EE. The remaining parts of this document are also part of the American National Standard and are identical to IEC 60601-2-4.

This is a preview edition of an AAMI guidance document and is intended to allow potential purchasers to evaluate the content of the document before making a purchasing decision. IEC 60601-2-4:2002 was developed by IEC Subcommittee (SC) 62D, Electromedical Equipment, which is administered by AAMI on behalf of the International Electrotechnical Commission (IEC), a worldwide organization for standardization. As previously noted, AAMI standards (as well as a previous edition of the IEC document) served as the basis of the international standard. In addition, the AAMI Defibrillator Committee, working as the U.S. Technical Advisory sub-Group for IEC SC 62D/WG2, was responsible for developing U.S. consensus on the international standard and otherwise participated in the drafting of that document. For a complete copy of this AAMI document, contact AAMI at (800) 332-2264, ext. 217

The concepts incorporated in this standard should not be considered inflexible or static. This standard, like any other, must be reviewed and updated periodically to assimilate progressive technological developments. To remain relevant, it must be modified as technological advances are made and as new data comes to light.

This standard reflects the conscientious efforts of concerned health care professionals and medical device manufacturers to develop a standard for those performance levels that can be reasonably achieved at this time.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Technical Programs, AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

NOTE—This background does not contain provisions of the American National Standard *Medical electrical equipment—Part 2-4: Particular requirements for the safety of cardiac defibrillators (including automated external defibrillators)* (ANSI/AAMI DF80:2003), but it does provide important information about the development and intended use of the document.
